

# Northumbria Research Link

Citation: Hall, Andrew, Mundell, Victoria, Blanco-Andujar, Cristina, Bencsik, Martin, McHale, Glen, Newton, Michael and Cave, Gareth (2010) Towards MRI microarrays. Chemical Communications, 46 (14). pp. 2420-2422. ISSN 0022-4936

Published by: UNSPECIFIED

URL:

This version was downloaded from Northumbria Research Link: <http://northumbria-test.eprints-hosting.org/id/eprint/47381/>

Northumbria University has developed Northumbria Research Link (NRL) to enable users to access the University's research output. Copyright © and moral rights for items on NRL are retained by the individual author(s) and/or other copyright owners. Single copies of full items can be reproduced, displayed or performed, and given to third parties in any format or medium for personal research or study, educational, or not-for-profit purposes without prior permission or charge, provided the authors, title and full bibliographic details are given, as well as a hyperlink and/or URL to the original metadata page. The content must not be changed in any way. Full items must not be sold commercially in any format or medium without formal permission of the copyright holder. The full policy is available online: <http://nrl.northumbria.ac.uk/policies.html>

This document may differ from the final, published version of the research and has been made available online in accordance with publisher policies. To read and/or cite from the published version of the research, please visit the publisher's website (a subscription may be required.)



UniversityLibrary



**Northumbria**  
**University**  
NEWCASTLE

# Northumbria Research Link

Citation: Hall, Andrew, Mundell, Victoria, Blanco-Andujar, Cristina, Bencsik, Martin, McHale, Glen, Newton, Michael and Cave, Gareth (2010) Towards MRI microarrays. *Chemical Communications*, 46 (14). pp. 2420-2422. ISSN 1359-7345

Published by: Royal Society of Chemistry

URL: <http://dx.doi.org/10.1039/B925020D> <<http://dx.doi.org/10.1039/B925020D>>

This version was downloaded from Northumbria Research Link:  
<http://nrl.northumbria.ac.uk/5178/>

Northumbria University has developed Northumbria Research Link (NRL) to enable users to access the University's research output. Copyright © and moral rights for items on NRL are retained by the individual author(s) and/or other copyright owners. Single copies of full items can be reproduced, displayed or performed, and given to third parties in any format or medium for personal research or study, educational, or not-for-profit purposes without prior permission or charge, provided the authors, title and full bibliographic details are given, as well as a hyperlink and/or URL to the original metadata page. The content must not be changed in any way. Full items must not be sold commercially in any format or medium without formal permission of the copyright holder. The full policy is available online: <http://nrl.northumbria.ac.uk/policies.html>

This document may differ from the final, published version of the research and has been made available online in accordance with publisher policies. To read and/or cite from the published version of the research, please visit the publisher's website (a subscription may be required.)

[www.northumbria.ac.uk/nrl](http://www.northumbria.ac.uk/nrl)




Electronic Supporting Information

## Towards MRI microarrays


Andrew Hall, Victoria J. Mundell, Cristina Blanco-Andujar, Martin Bencsik, Glen McHale, Michael I. Newton and Gareth W. V. Cave<sup>a</sup>

Slide 1.




The array was made as described in the communication; however, the two separate lines of SPIO spots, the bottom line with spot size  $\sim 2$  mm and the top line with spot size  $\sim 1$  mm are all positive and imaged in water. It was imaged with increasing spatial resolution of (a)  $1370 \times 719 \mu\text{m}^2$ , (b)  $978 \times 511 \mu\text{m}^2$ , (c)  $688 \times 359 \mu\text{m}^2$  and (d)  $344 \times 180 \mu\text{m}^2$ . Bright pixels in close proximity to dark pixels are seen at the spots, which is the signature of MRI 'susceptibility effect' artefact. Increasing spatial resolution requires longer imaging time ( $= 28$  s for (a),  $56$  s for (b) and (c),  $224$  s for (d)), but allows enhanced identification of the smallest spots. The slice thickness is 1 mm for all images and the echo time value was individually tailored to reach a good compromise of SNR to contrast.

Slide 2.



The slide was made by laser etching a 5 x 15 array and spotting with cDNA (all the same). cDNA was then spotted onto the spots directly and allowed to dry. This was subsequently washed before imaging in water. MRI details: RARE sequence, RARE factor = number of lines = 64,  $T_E = 30$  ms,  $T_E^{\text{eff}} = 975$  ms, 8 averages, spatial resolution =  $1172 \times 547 \mu\text{m}^2$ , slice thickness = 0.5 mm, imaging time = 15.5 s.

Slide 3.



The slide was made by laser etching “M-RAY” the glass slide, and then immobilising cDNA into the etch surface before hybridising with cDNA (as described in the communication) and imaging in water. MRI details: RARE sequence, RARE factor = number of lines = 128,  $T_E = 30$  ms,  $T_E^{\text{eff}} = 1726$  ms, 16 averages, spatial resolution =  $781 \times 391 \mu\text{m}^2$ , slice thickness = 1 mm, imaging time = 62 s.